

BID BULLETIN NO. BAC-G-PTRE-2022-01

Rebidding for the Supply, Delivery, and Installation of Physical Therapy and Rehabilitation Equipment for the NAS – Phase 1 at New Clark City.

We are pleased to provide further clarifications on the following eligibility requirements for the **Rebidding for the Supply, Delivery and Installation of Physical Therapy and Rehabilitation Equipment for the NAS – Phase 1 at New Clark City:**

Queries		Clarification
1.	The required warranty, say 12 months, will it start from delivery or acceptance?	The warranty starts upon the issuance of a Certificate of Acceptance.
2.	May we clarify if this procurement is a lot bidding?	Yes, the bidder will bid for the entire lot.
3.	For the Anti-Gravity Treadmill, is the harness mounted to the ceiling or is the harness attached to the treadmill?	This is a standalone treadmill, it is not mounted on the ceiling.
4.	For Vinyl Coated Dumbbells, is it possible to provide dumbbells in pounds instead of kg?	The stated specification is a minimum requirement, the bidders can submit higher specs, if converted in lbs. The dumbbells must have a weight equivalent to the required weight in kilograms.
5.	For Parallel Bars, is it required that it is floor/platform mounted?	Yes.
6.	For Hydrocollator, is there a minimum volume for this?	The minimum volume should be large enough to accommodate all hot moist packs indicated in this lot.
7.	Can we request a 2 weeks extension for the submission of bids?	No, we can not extend the date of submission of the bids. This is one of the top priority projects since some NAS classes are now ongoing in other temporary facilities. The delivery period of the equipment has been given much consideration already.
8.	Please clarify the cost of bidding documents for this project.	Pursuant to Annex 8 of the RIRR of RA 9184, since the ABC for this project is Php6,639,000.00 , the cost of its bidding documents is Php10,000.00 .
9.	For Item 2. The specifications show two different track lengths while the quantity stated is 1 set (comes with a rail track, sport sling harness set, sling roller unit, pulley, and ropes). May we rephrase the specifications of 7.5ft minimum length and 10ft maximum length? May we also offer its equivalent in metric?	It is acceptable to indicate the metric equivalent as long as its conversion measurement is equivalent to 7.5ft to 10ft length.
10.	For Item 3, regarding the programmable position of 300, is it for the patient database or custom protocol? It so happened that our product has only 25 customizable protocols on the machine itself, but it has unlimited programmable positions on the patient database and protocols when connected to a windows 10 PC thru Bluetooth.	The 25 customizable protocol machine is acceptable provided that the supplier will provide support on programming the protocols using their own Bluetooth enabled Windows 10 PC when needed and free of charge.





11. For Item 8, can you consider the following?	YES
a. Physio ball, Ø 80-85 mm	
b. Physio ball, Ø 95-100 mm	
	YES. For the height, we will conditionally allow the requested height allowance but since it is higher than our standard, the supplier should provide a stable & sturdy footstool.
 13. For Item 16. Can you consider the following? Choice of height range - 45cm-59cm or 54cm-(73-74cm) Foot Support Ring fitted as standard on High Model MC6163 (specific brand model) Seat Pad Dimensions: (350-400mm)x(350-400mm)x 90mm 	Item 16 is a hydrocollator. If you are referring to Item 14, all adjustments will be conditionally allowed, provided that it can accommodate the 5 th to 95 th percentile of Filipino standard anthropometric measurements.
 14. For Item 16. Can you consider the following: 220V AC, 60Hz, 1,500W-2900W Max. Volume: 80-88 Liters Internal Size: (49-50)x38x(45-46.5)cm (WxDxH) External Size: (56-58.5)x(45.5-47)x(71-81cm) 	All requested adjustments will be allowed except the 2900W which will be conditionally allowed provided that the supplier confirms that the NAS PT Facility electrical outlet and capacity can accommodate this specification.
	 All requested adjustment will be conditionally allowed provided that the following hot packs purposes are accounted: a) Cervical Application b) Whole Back Application (Oversize) c) One side back application (Standard) d) Knee Application (With hole in the middle)
16. For Item 16. Can you consider:	YES
(230-280)x(130-180mm) Gel Type	
17. For Item 20. Can you consider the	YES
following:	
Size: (33-42.5) * (27-28) * (19-23.5cm)	
Temperature range: 45-65 degrees Celsius	



Reminders				
Single Largest Completed Contract (SLCC)	The Bidder must have completed a single contract that is similar to the project, equivalent to at least fifty percent (50%) of the ABC; or at least two (2) similar completed contracts with the aggregate amount equivalent to at least fifty percent (50%) of the ABC and the largest of these similar contracts must be equivalent to at least half of the percentage (25%) of the ABC. A similar contract shall have been completed within ten (10) years prior to the date of submission and receipt of a bid. For this purpose, contracts similar to the Project shall be			
	 Contracts for the Supply, Delivery, and Installation of Physical Therapy & Rehabilitation Equipment; and/or Contracts for the Supply, Delivery, and Installation of Sports Equipment. 			
Financial Component				
Dreduct Drechurg	for the Project			
Product Brochure OR Manufacturer's Certificate	It is <u>required</u> that the Bidders submit <u>Product Brochures and/or other related</u> <u>documents</u> containing the description and specifications of each item to be offered and/ <u>OR Certificate from the Manufacturer</u> , certifying compliance in all the specifications of each item to be offered as part of their technical envelope during the opening of bids.			

Moreover, for clarification purposes, please take note of the following schedule:

Activities	Date/Schedule
Deadline for Request for Clarification, if any	19 July 2022 (Tuesday)
Issuance of Bid Bulletin, if any	25 July 2022 (Monday)
Deadline for Submission of the ff: • Eligibility Requirements • Financial Proposal	9:00 AM, 02 August 2022 (Tuesday)
Opening of Bids	11:00 AM, 02 August 2022 (Tuesday)

The above changes further amend the bidding documents, accordingly. The Opening of Bids will be held face-to-face at the BCDA Board Room, 2/F Bonifacio Technology Center, 31st corner 2nd Avenue, Bonifacio Global City, Taguig City, with video coverage.

Bidders may choose to attend the Opening of Bids in person or online via zoom.

For those who are attending in person, the following guidelines must be followed:

- A maximum of two (2) persons per participating bidder are allowed to attend the Opening of Bids and will be required to follow the BCDA Health Protocol for Visitors;
- The representative is required to present his/her vaccination card before he/she can be allowed inside the BCDA office;
- Compliance to social distancing, wearing of face masks, and body temperature screening inside BCDA office must be observed; and





• Observers/representatives who show signs of COVID-19 related symptoms such as cough, flu, fever, high body temperature, and sneezing are advised not to proceed to the BCDA Office since they will not be allowed to enter the BCDA premises.

This Notice is being issued pursuant to Section 22.5.2 and 22.5.3 of the 2016 Revised Implementing Rules and Regulations of Republic Act 9184.

Issued on 25 July 2022 (Monday).

BIDS AND AWARDS COMMITTEE FOR GOODS

By:

JA M. BINGCA

